

PRODUCT BULLETIN

Product(s):	Fabius Tiro
Bulletin Number:	154
Release Date:	September 29, 2003
Pages:	1

FOR INTERNATIONAL RELEASE ONLY!

Subject: Gas Inlet Shield

Effective October 2003, all new Fabius Tiro anesthesia systems will be shipped with a Gas Inlet Shield. For installed units in the field, a service kit will be available under part number 4118541. Availability date of the service kits will be provided by service.

This shield is necessary for directing any possible oxygen leaks away from the Desflurane vaporizer and limiting the oxygen enrichment in the vaporizer area.

The Gas Inlet Shield is mandatory on all Fabius Tiro machines and must be installed on any units currently in the field that do not have it already.

Please contact Christine MacMillan for any questions at ext. 2303 or macmillanc@draegermed.com

